



REVISION	
05-22-01	UPDATED SCHEMATIC WITH NEW EL-3013 BOX TO HOUSE THE SOLID-STATE RELAYS
07-11-01	ADDED WIRE DETAILS
08-22-01	CORRECTED CONNECTOR DETAILS
08-23-01	REPLACED 2 SOLID STATE RELAYS WITH 1 CONVENTIONAL RELAY
08-30-01	REMOVED AC WIRES GOING THROUGH BOX AT J4,D&E TO PROTEUS FLOW SWITCH. CORRECTED PIN NUMBERS ON RELAY.
09-17-01	ADDED FLOW SWITCH CALIBRATION AND CONTACT INFO AND CORRECTED K1 PIN NUMBERS.

UNIVERSITY OF CALIFORNIA LICK OBSERVATORY		TV AC POWER INTERRUPT BOX DEIMOS SPECTROGRAPH	
DES'N BY: B. Alcott	ORIGIN DATE: 01-02-01	DWG. NO.	NUM. 1 OF 1
Last SN	MODIFY DATE: 09-17-01	EL-3013	
PATH: DEIMOS/TVINTERRUPT	REVA		